

VOLATILE ORGANIC COMPOUNDS
EMISSIONS COMPLIANCE TESTING
AT PAXFORD COMPOSITES LTD IN MARCH 2010

FOR: Paxford Composites Ltd
2-4 Redwongs Way
Huntingdon
Cambridgeshire
PE29 7HB

FAO: Mr Neil Search

Work By: K C Blakley and M R Ellison

Reference:105435\QE8400\PX02MAR10

Page 1 of 22

Date of issue: NPL Authorised Signature

Name:

MCerts Level 2 Approver

Signed: for Managing Director

Name: Kevin Blakley

Signature:

MCERTs No: MM-03-317

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Part 1: - Executive Summary –Compliance Report		
Process Operator Paxford Composites Ltd		LA Permit No B01/02
		Site Specific Protocol Reference PX02MAR10\SSP
Address Paxford Composites Ltd 2-4 Redwongs Way Huntingdon Cambridgeshire PE29 7HB		Contact Neil Search
		Tel No 01480 453537
		Email neil.search@paaxfordcomposites.co.uk
Tests carried out	Compliance VOCs of 3 Spray Booths	Dates tests carried out 31st March 2010
Testing laboratory National Physical Laboratory		UKAS Accreditation No 0002
Address Hampton Road Teddington TW11 OLV		Contact Kevin Blakley
		Tel No 020 8943 6118
		Email kevin.blakley@npl.co.uk
Species to be monitored	Volatile Organic Compounds	
Emissions Limit Values (ELV)	Volatile Organic Compounds - 150 mg m⁻³ (expressed as carbon)	
Compliance with standards	Yes	
Deviations from standards		No
Corrective actions required		No
Test Team	Kevin Blakley and Matthew Ellison	
Full report reference number	PX02MAR10	
Summary report submitted by Kevin Blakley	Signature	MCERTs ID No: - Level 2, TE1,2,3,4

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

CONTENTS LIST

	Page
PART 1:EXECUTIVE SUMMARY	
PART 1 COVER SHEET	2
1.1 MONITORING OBJECTIVES	4
1.2 VOC MONITORING RESULTS	4
1.3. PLANT AND EQUIPMENT OPERATING INFORMATION	4
1.4 MONITORING DEVIATIONS	5
1.5 CONCLUSIONS	5
1.6 BIBLIOGRAPHY	5
PART 2: SUPPORTING INFORMATION	
PART 2 COVER SHEET	6
2.1 APPENDIX 1	8
2.1.1 Emission Testing Personnel Information	8
2.1.2 Test Techniques and Protocols	8
2.2 APPENDIX 2	
2.2.1 Sample Points Details	9
2.2.2 VOC 1 Minute Averaged Graphs	11
2.2.3 Gas Measurements	15
2.2.4 Calculations Used in Reporting Results	17
2.2.5 Uncertainty Calculations	18

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.1 MONITORING OBJECTIVES

NPL were awarded a contract by Paxford Composites Ltd to conduct emissions compliance testing at their facility located in Huntingdon, Cambridgeshire. Work was conducted on the 18th February to make the necessary measurements from their spray painting enclosures, known as Spray Booth 1 (SB1), Spray Booth 2 (SB2) and Spray Booth 3 (SB3).

The requirements of the contract were to quantify volatile organic compounds (VOC's) from the sources described above over a 30-minute period.

VOC concentrations are expressed as carbon at reference conditions of 273.15K, 101.325 kPa on a wet gas basis.

1.2 VOC MONITORING RESULTS

Field	Units			
Stack I.D.		SB1	SB2	SB3
Date		31/03/2010	31/03/2010	31/03/2010
Sample Period (BST)	From hh:mm	10:15	12:45	14:00
	To hh:mm	10:45	13:15	14:30
Sample Duration	min	30	30	30
30-minute mean Concentration for Period	mg m ⁻³ , Ref. Conditions	20.5	30.9	25.3
Expanded Uncertainty	mg m ⁻³ , 95% conf. k=2	+/-1.3	+/-1.4	+/-1.4
Emission Limit Value (ELV)	mg m ⁻³ , Ref. Conditions	150	150	150
Percentage of ELV	%	14	21	17
Reference Conditions	273.15K, 101.325 kPa, Wet Gas Basis			

Notes: -

The VOC (mg/m³) results above are expressed as Carbon and these were calculated using the prescribed method described in the Environmental Agency Technical Guidance Note M16 for Volatile Organic Compounds.

1.3 PLANT AND EQUIPMENT OPERATING INFORMATION

1.3.1 Paxford Composites Spray Booths 1-3

The spray booths at Paxford Composites consist of three sealed rooms, approximately the size of a large household car garages. They are used as batch processes and air is pumped into the booths from outside the building and the spray painting process is carried out manually using 2 to 3 skilled workers. The air inside the booth can be heated if required for curing treatment of the components. Owing to the business requirements, many layers of paint are required to achieve the smooth finish. As a result, a typical single layer spray time would normally be approximately 20 minutes but depends on the size and type of item being sprayed. This process is then repeated after the item has been brushed down and ready for a further coating. The paint filled air inside the booths are removed via ceiling filters (which covers the entire surface areas of the ceilings) to remove particulate matter. The air is then exhausted to atmosphere.

On each spray booth, the sample position was downstream of the particulate filters. The sample points were 10mm holes, suitable for monitoring VOCs to the relevant CEN Standard, See Figures 1-3.

At the time of monitoring, spray painting workers were asked to carry out their normal task in order to create representative conditions of routine activity within the spray booths. It should be noted that at the time of monitoring, there were no items being physically sprayed, therefore all of the paint would have been expelled through the exhaust ducting rather than coating surfaces of the items. It was the site's opinion that this may lead to higher levels of VOCs being emitted than would usually be seen.

1.4 MONITORING DEVIATIONS

The testing was fully compliant with the relative standard BS EN13526: 2002. See Appendix 1 for Test Techniques and Protocols.

1.5 CONCLUSIONS

Concentrations of volatile organic compounds, measured from the three spray booths at Paxford Composites Ltd, were carried out on the 31st March 2010.

Particulates monitoring could not be undertaken as required in the sites permit due to the size of the ports available. Therefore, it is recommended the site fit suitable ports to allow the particulate monitoring to be undertaken as described in BS EN15259: 2007.

1.6 REFERENCES

1. Guidance on Assessing Measurement Uncertainty in Stack Emissions Monitoring, by Pullen J and Robinson R, Source Testing Association, Quality Guidance Note QGN1.

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Part 2: - Supporting Information–Compliance Report			
Process Operator Paxford Composites Ltd		LA Permit No B01/02	
		Site Specific Protocol Reference PX02MAR10\SSP	
Address Paxford Composites Ltd 2-4 Redwongs Way Huntingdon Cambridgeshire PE29 7HB		Contact Neil Search	
		Tel No 01480 453537	
		Email neil.search@paxfordcomposites.co.uk	
Tests carried out	Compliance VOCs of 3 Spray Booths	Dates tests carried out 31st March 2010	
Testing laboratory National Physical Laboratory		UKAS Accreditation No 0002	
Address Hampton Road Teddington TW11 OLW		Contact Kevin Blakley	
		Tel No 020 8943 6118	
		Email kevin.blakley@npl.co.uk	
Species to be monitored		Volatile Organic Compounds	
Emissions Limit Values (ELV)	Volatile Organic Compounds - 150 mg m⁻³ (expressed as carbon)		
Compliance with standards		Yes	
Deviations from standards			No
Corrective actions required			No
Test Team	Kevin Blakley and Matthew Ellison		
Full report reference number		PX02MAR10	
Summary report submitted by Kevin Blakley		Signature	MCERTs ID No: - Level 2, TE1,2,3,4

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

CONTENTS LIST

	Page
PART 2: SUPPORTING INFORMATION	
PART 2 COVER SHEET	6
2.1 APPENDIX 1	
2.1.1 Emission Testing Personnel Information	8
2.1.2 Test Techniques and Protocols	8
2.2 APPENDIX 2	
2.2.1 Sample Points Details	9
2.2.2 VOC 1 Minute Averaged Graphs	11
2.2.3 Gas Measurements	15
2.2.4 Calculations Used in Reporting Results	17
2.2.5 Uncertainty Calculations	18

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

2.1 APPENDIX 1

2.1.1 Emission Testing Personnel Information

NAME	NPL Position	MCerts I.D. No.	Level/Endorsements	Function
Matthew Ellison	RS	MM-05-682	Level 2, TE1, TE2, TE3, TE4	Team Leader
Kevin Blakley	SRS	MM-03-317	Level 2, TE1, TE2, TE3, TE4	Team Leader

2.1.2 Test Techniques and Protocols

Date of Tests	31 st March 2010
Main Pollutants Measured	1. Volatile Organic Compounds (VOCs)
Test Methods	1. Flame Ionisation Detector, Sick-Maihak. Serial No. AS0202 to BS EN 13526:2002. MCertified Instrument No – Sira MC 040037/02
Procedure	1. QPAS B 538:- STACK GAS SAMPLING FOR CARBON MONOXIDE, CARBON DIOXIDE, OXYGEN, SULPHUR DIOXIDE, VOLATILE ORGANIC COMPOUNDS AND NITROGEN OXIDES USING EXTRACTIVE INSTRUMENTAL TECHNIQUES.

2.2 APPENDIX 2

2.2.1 Sample Point Details

Figure 1. Spray Booth 1 Sample Position



Figure 2. Spray Booth 2 Sample Position






Figure 3. Spray Booth 3 Sample Position



2.2.2 VOC 1 Minute Averaged Graphs


NATIONAL PHYSICAL LABORATORY
Continuation Sheet

2.2.2.1 Figure 4. Spray Booth 1. 1-Minute Averages of Volatile Organic Compounds (VOC).
Expressed as mg m^{-3} (Carbon) at 273K, 101.3kPa, Wet Gas Basis




NATIONAL PHYSICAL LABORATORY
Continuation Sheet

2.2.2.2 Figure 5. Spray Booth 2. 1-Minute Averages of Volatile Organic Compounds (VOC).
Expressed as mg m⁻³ (Carbon) at 273K, 101.3kPa, Wet Gas Basis



NATIONAL PHYSICAL LABORATORY
Continuation Sheet

2.2.2.3 Figure 6. Spray Booth 3. 1-Minute Averages of Volatile Organic Compounds (VOC).
Expressed as mg m^{-3} (Carbon) at 273K, 101.3kPa, Wet Gas Basis



2.2.3 Gas Measurements

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Calibration Log

Instrument SN: A5-0202		Type			Job No: PAX02MAR10			Page No:	
Mobile Lab: VOC		.02			Client Site: PAXFORD COMPOSITES			2 of 3	
			NPL		Other Equipment: NONE			A3	
Gas Cyl. No.:	115985	Gas Cyl Conc.:	51.4	<i>PACKAGE 4 vol ppm</i>	Date:	31/3/10	Instrument Range:	0-100 ppm	
Cal Type	Time (Start and Finish)	Initial Reading (V)	Initial Gain/Setting (if app.)	Final Reading (V)	Final Gain/Setting (if app.)	System(S) / Direct(D)	Total Regulator Pressure (bar/Pa) delete as app.)	Comment	Signature
Check Zero	14:45					D		Zero drift: <2%	WBS
Zero Adj	14:46	0.064	4.75	0.007	4.65	D	210	Span Drift: <2%	WBS
Check Span	14:47					D	60	Response Time: 20 sec	WBS
Span Adj	14:48	5.109	2.71	5.142	2.73	D		Span Value: 514%	WBS
Check Zero	14:49	0.002	4.65			D	210		WBS
Check Zero	15:22	0.009	4.65			D	210		WBS
Check Span	15:23	5.062	2.73			D	60		WBS
Zero Adj									
Span Adj									
Check Zero	15:25	0.005	2.73			D	210		WBS

Instrument SN:		Type							
Mobile Lab:		CO2							
			NPL						
Gas Cyl. No.:		Gas Cyl Conc.:		% vol	Date:		Instrument Range:		
Cal Type	Time (Start and Finish)	Initial Reading	Initial Gain/Setting (if app.)	Final Reading	Final Gain/Setting (if app.)	System(S) / Direct(D)	Total Regulator Pressure (bar/Pa) delete as app.)	Comment	Signature
Check Zero								Zero drift: ____%	
Zero Adj								Span Drift: ____%	
Check Span								Response Time: ____ sec	
Span Adj								Span Value: ____%	
Check Zero									
Check Zero									
Check Span									
Zero Adj									
Span Adj									
Check Zero									

IF REGULATOR PRESSURE IS LESS THAN 500 PSI (35 bar) CONTACT SUPPLIER IMMEDIATELY

MCERTS LEVEL2 with TE4
MCERTS ID: **MM-03-317**

Name: **K. BUCKLEY**

NPL Ltd: 15 OCT 2009
Author: KCB

Revision 3
QPAS B 638

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Calibration Log

Instrument SN:		Type	Job No:				Page No:		
A5-0202			PX02MAR10				1 of 3		
Mobile Lab: TRANSIT		VOC	Client Site: Payford Composites - A1						
		NPL	Other Equipment:						
Gas Cyl. No.	115985	Gas Cyl Conc.	51.4	ppm PROPANE	Date:	31/03/10	Instrument Range: 0-100 ppm		
Cal Type	Time (Start and Finish)	Initial Reading (V)	Initial Gain/Setting (if app.)	Final Reading (V)	Final Gain/Setting (if app.)	System(S) / Direct(D)	Total Regulator Pressure (bar/Psi delete as app.)	Comment	Signature
Check Zero	10:17	0.248	4.80	0.002	4.84	D	3000psi	Zero drift: <2%	WB
Zero Adj	10:18							Span Drift: <2%	
Check Span	10:21	5.380	2.68	5.148	2.58	D	60Bar	Response Time: 20 sec	WB
Span Adj	10:23							Span Value: 51.4 ppm	
Check Zero	10:25	4.64	0.003			D	3000psi		WB
Check Zero	12:20	0.001	4.64			D	3000psi		WB
Check Span	12:23	5.192	2.58			D	60Bar		WB
Zero Adj									
Span Adj									
Check Zero	12:24	0.002	4.64			D	3000psi		WB
Instrument SN:		Type	Job No:						
A5-0202			PX02MAR10						
Mobile Lab:		VOC	Client Site: Payford Composites - A1						
		NPL	Other Equipment:						
Gas Cyl. No.	115985	Gas Cyl Conc.	51.4	ppm PROPANE	Date:	31/3/10	Instrument Range: 0-100 ppm		
Cal Type	Time (Start and Finish)	Initial Reading (V)	Initial Gain/Setting (if app.)	Final Reading (V)	Final Gain/Setting (if app.)	System(S) / Direct(D)	Total Regulator Pressure (bar/Psi delete as app.)	Comment	Signature
Check Zero	11:46	-0.3	-	-0.01	-	D	3000psi	Zero drift: <2%	ME
Zero Adj	11:48							Span Drift: <2%	
Check Span	11:52	4.919	-	5.138	-	D	60Bar	Response Time: 20 sec	ME
Span Adj	11:54							Span Value: 51.4 ppm	
Check Zero	11:57	0.009	-			D	3000psi		ME
Check Zero	14:15	0.007	-			D	3000psi		ME
Check Span	14:17	5.129	-			D	60Bar		ME
Zero Adj									
Span Adj									
Check Zero	14:19	0.003	-			D	3000psi		ME

IF REGULATOR PRESSURE IS LESS THAN 500 PSI (35 bar) CONTACT SUPPLIER IMMEDIATELY

MCERTS LEVEL2 with TE4
MCERTS ID: MM-03-317

Name: K. BLAKLEY

NPL Ltd: 15 OCT 2009
Author: KCB

Revision 3
QPAS B 538

2.2.4 Calculations Used in Reporting Results

The following equation can be used to convert propane volume concentrations to total organic carbon mass concentrations:

$$C_m = C_v \frac{3 \times M_c}{V_m} \text{ mg/m}^3$$

where:-

C_m is the TOC concentration in milligrams per cubic metre (273 K;1013 hPa).

C_v is the volume concentration of propane in ppm (by volume).

M_c is the molar mass of carbon (=12 g/mole).

V_M is the molar volume (=22,4 l/mole).

2.2.5 Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 1

Limit value	150	mg.m ⁻³ (corrected) Carbon	Gas	Propane	
Measured concentration	12.77	ppm	Full Scale	100	ppm
Measured concentration	20.50	mg.m ⁻³ (273K, 101.3kPa) Carbon	Cal gas conc	51.4	ppm
			Conversion	0.62305296	
			Full Scale	62.30529595	mg.m ⁻³ (Carbon)
			Cal gas conc	32.02492212	mg.m ⁻³ (Carbon)

Correction for reference conditions					
	ref	O2, %	Moisture, %	Pressure, KPa	Temperature, K
	ref	21.00	0.00	101.30	273.00
	measured	21.00	0.00	101.30	273.00
Factors		1.00	1.00	1.00	1.00
Correction Factor		1.00			

Performance characteristics	Value		specification
Response time	20	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0	% of value/kPa	0.750
ambient temperature dependence	0.025	% full scale/10K	0.300
NH3 (20 mg/m3)	0	mg/m3	
CO2 (15%)	0	% by vol	
H2O (30%)	0.0	% by vol	4.000
dependence on voltage	1.05	% full scale/10V	2%fs/10V
losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m3
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH3 range	0	0	0
CO2 range	0	15	0
H2O range	0	0	0
Instrument Voltage Rating			110
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions			
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		U _{lo}	for mean use rep at span
Standard deviation of repeatability at span level		U _{rs}	for mean 0.02
Lack of fit		U _{fit}	-0.53
Drift		U _{odr}	0.00
volume or pressure flow dependence		U _{spres}	0.00
atmospheric pressure dependence		U _{apres}	0.00
ambient temperature dependence		U _{temp}	0.00
NH3 (20 mg/m3)		U _{interf}	0.00
CO2 (15%)			0.00
H2O (30%)			0.00
Dependence on voltage		U _{volt}	0.33
losses in the line (leak)		U _{leak}	0.00
Uncertainty of calibration gas		U _{calib}	0.24

Use largest negative or positive interferent effect		
	0	0.00
	0	0.00
	0	0.00
	0	0.00
Interference uncertainty		0.00

Measurement uncertainty	Result		
Combined uncertainty	20.50	mg/m ³	
Expanded uncertainty	0.67	mg/m ³	
Expanded uncertainty	1.33	mg/m ³	
Expanded uncertainty	1.33	mg.m-3 (corrected)	
Expanded uncertainty expressed with a level of confidence of 95%	0.89	% ELV	
Expanded uncertainty expressed with a level of confidence of 95%	1.33	mg.m ⁻³	

Note: Enter values into green boxes
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R. Robinson, NPL

Reference: 105435\QE8400\PX02MAR10
Permit No: B01/02
Checked by:

Version 1

NATIONAL PHYSICAL LABORATORY

Continuation Sheet

Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 2

Limit value	150	mg.m ⁻³ (corrected) Carbon	Gas	Propane
Measured concentration	19.25	ppm	Full Scale	100
Measured concentration	30.90	mg.m ⁻³ (273K, 101.3kPa) Carbon	Cal gas conc	51.4
			Conversion	0.62305296
			Full Scale	62.30529595
			Cal gas conc	32.02492212

Correction for reference conditions					
	ref	O2, %	Moisture, %	Pressure, KPa	Temperature, K
		21.00	0.00	101.30	273.00
	measured	21.00	0.00	101.30	273.00
Factors		1.00	1.00	1.00	1.00
Correction Factor		1.00			

Performance characteristics	Value		specification
Response time	20	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0	% of value/kPa	0.750
ambient temperature dependence	0.025	% full scale/10K	0.300
NH3 (20 mg/m3)	0	mg/m3	
CO2 (15%)	0	% by vol	
H2O (30%)	0.0	% by vol	4.000
dependence on voltage	1.05	% full scale/10V	2%fs/10V
losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m3
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH3 range	0	0	0
CO2 range	0	15	0
H2O range	0	0	0
Instrument Voltage Rating	110		
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions			
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		U _{r0}	for mean
Standard deviation of repeatability at span level		U _{rs}	use rep at span
Lack of fit		U _{fit}	-0.53
Drift		U _{odr}	0.00
volume or pressure flow dependence		U _{spres}	0.00
atmospheric pressure dependence		U _{apres}	0.00
ambient temperature dependence		U _{temp}	0.00
NH3 (20 mg/m3)		U _{interf}	0.00
CO2 (15%)			0.00
H2O (30%)			0.00
Dependence on voltage		U _{volt}	0.33
losses in the line (leak)		U _{leak}	0.00
Uncertainty of calibration gas		U _{calib}	0.36

Use largest negative or positive interferent effect	
0	0.00
0	0.00
0	0.00
0	0.00
Interference uncertainty	0.00

Measurement uncertainty	Result		
Combined uncertainty		0.72	mg/m ³
Expanded uncertainty	k = 2	1.43	mg/m ³
Expanded uncertainty	expressed with a level of confidence of 95%	0.96	% ELV
Expanded uncertainty	expressed with a level of confidence of 95%	1.43	mg.m ⁻³

Note: Enter values into green boxes
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R Robinson, NPL

Reference: 105435\QE8400\PX02MAR10

Permit No: B01/02

Checked by:

Version 1

NATIONAL PHYSICAL LABORATORY

Continuation Sheet

Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 3

Limit value	150	mg.m ⁻³ (corrected) Carbon	Gas	Propane
Measured concentration	15.76	ppm	Full Scale	100
Measured concentration	25.30	mg.m ⁻³ (273K, 101.3kPa) Carbon	Cal gas conc	51.4
			Conversion	0.62305296
			Full Scale	62.30529595
			Cal gas conc	32.02492212

Correction for reference conditions					
	ref	O2, %	Moisture, %	Pressure, KPa	Temperature, K
		21.00	0.00	101.30	273.00
	measured	21.00	0.00	101.30	273.00
Factors		1.00	1.00	1.00	1.00
Correction Factor		1.00			

Performance characteristics	Value		specification
Response time	20	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0	% of value/kPa	0.750
ambient temperature dependence	0.025	% full scale/10K	0.300
NH3 (20 mg/m3)	0	mg/m3	
CO2 (15%)	0	% by vol	
H2O (30%)	0.0	% by vol	4.000
dependence on voltage	1.05	% full scale/10V	2%fs/10V
losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m3
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH3 range	0	0	0
CO2 range	0	15	0
H2O range	0	0	0
Instrument Voltage Rating	110		
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions			
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		U _{r0}	for mean
Standard deviation of repeatability at span level		U _{rs}	use rep at span
Lack of fit		U _{fit}	for mean
Drift		U _{odr}	-0.53
volume or pressure flow dependence		U _{spres}	0.00
atmospheric pressure dependence		U _{apres}	0.00
ambient temperature dependence		U _{temp}	0.00
NH3 (20 mg/m3)		U _{interf}	0.00
CO2 (15%)			0.00
H2O (30%)			0.00
Dependence on voltage		U _{volt}	0.33
losses in the line (leak)		U _{leak}	0.00
Uncertainty of calibration gas		U _{calib}	0.29

Use largest negative or positive interferent effect	
0	0.00
0	0.00
0	0.00
0	0.00
Interference uncertainty	0.00

Measurement uncertainty	Result		
Combined uncertainty		25.30	mg/m ³
Expanded uncertainty		0.69	mg/m ³
Expanded uncertainty	k = 2	1.37	mg/m ³
Expanded uncertainty			
Expanded uncertainty		1.37	mg.m-3 (corrected)
Expanded uncertainty	expressed with a level of confidence of 95%	0.92	% ELV
Expanded uncertainty	expressed with a level of confidence of 95%	1.37	mg.m ⁻³

Note: Enter values into green boxes
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R Robinson, NPL

Reference: 105435\QE8400\PX02MAR10
Permit No: B01/02
Checked by:

Version 1