

VOLATILE ORGANIC COMPOUNDS  
EMISSIONS COMPLIANCE TESTING  
AT PAXFORD COMPOSITES LTD IN FEBRUARY 2009

FOR: Paxford Composites Ltd  
2-4 Redwongs Way  
Huntingdon  
Cambridgeshire  
PE29 7HB

FAO: MR GRAHAME BLOXHAM

Work By: K C Blakley and R E Elliott

**Reference:**105435\QE8400\PX01FEB09

**Page 1 of 23**

**Date of issue:**

**Name: Mr R Robinson**

**(Authorised Signatory)**

**Signed:**

**for Managing Director**

**MCerts Level 2 Approver**

**Name:**

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

<b>Part 1: - Executive Summary – Compliance Report</b>			
Process Operator <b>Paxford Composites Ltd</b>		LA Permit No <b>B01/02</b>	
Address <b>Paxford Composites Ltd</b> <b>2-4 Redwongs Way</b> <b>Huntingdon</b> <b>Cambridgeshire</b> <b>PE29 7HB</b>		Contact <b>Grahame Bloxham</b>	
		Tel No <b>01480 453537</b>	
		Email <a href="mailto:Grahame.bloxham@paxfordcomposites.co.uk">Grahame.bloxham@paxfordcomposites.co.uk</a>	
Tests carried out	<b>Compliance VOCs of 3 Spray Booths</b>	Dates tests carried out <b>18<sup>th</sup> February 2008</b>	
Testing laboratory <b>National Physical Laboratory</b>		UKAS Accreditation No <b>0002</b>	
Address <b>Hampton Road</b> <b>Teddington</b> <b>TW11 OLW</b>		Contact <b>Kevin Blakley</b>	
		Tel No <b>020 8943 6118</b>	
		Email <a href="mailto:kevin.blakley@npl.co.uk">kevin.blakley@npl.co.uk</a>	
Species to be monitored		<b>Volatile Organic Compounds</b>	
Emissions Limit Values (ELV)		<b>Volatile Organic Compounds - 150 mg m<sup>-3</sup> (expressed as carbon)</b>	
Compliance with standards		<b>Yes</b>	
Deviations from standards			<b>No</b>
Corrective actions required			<b>No</b>
Test Team	<b>Kevin Blakley and Robert Elliott</b>		
Full report reference number		<b>PX01FEB09</b>	
Summary report submitted by  <b>Robert Elliott</b>		Signature	MCERTs ID No: -  <b>Level 2, TE1,2,3,4</b>

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**CONTENTS LIST**

	Page
<b>PART 1:EXECUTIVE SUMMARY</b>	
<b>PART 1 COVER SHEET</b>	2
1.1 <b>MONITORING OBJECTIVES</b>	4
1.2 <b>VOC MONITORING RESULTS</b>	4
1.3. <b>PLANT AND EQUIPMENT OPERATING INFORMATION</b>	4
1.4 <b>MONITORING DEVIATIONS</b>	5
1.5 <b>CONCLUSIONS</b>	5
1.6 <b>BIBLIOGRAPHY</b>	5
 <b>PART 2: SUPPORTING INFORMATION</b>	
<b>PART 2 COVER SHEET</b>	6
2.1 <b>APPENDIX 1</b>	8
2.1.1  Emission Testing Personnel Information	8
2.1.2  Test Techniques and Protocols	8
 2.2 <b>APPENDIX 2</b>	
2.2.1  Sample Points Details	9
2.2.2  VOC 1 Minute Averaged Graphs	11
2.2.3  Gas Measurements	15
2.2.4  Calculations Used in Reporting Results	18
2.2.5  Uncertainty Calculations	19

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### 1.1 MONITORING OBJECTIVES

NPL were awarded a contract by Paxford Composites Ltd to conduct emissions compliance testing at their facility located in Huntingdon, Cambridgeshire. Work was conducted on the 18<sup>th</sup> February to make the necessary measurements from their spray painting enclosures, known as Spray Booth 1 (SB1), Spray Booth 2 (SB2) and Spray Booth 3 (SB3).

The requirements of the contract were to quantify volatile organic compounds (VOC's) from the sources described above on a continuous basis.

VOC concentrations are expressed as carbon at reference conditions of 273.15K, 101.325 kPa on a wet gas basis.

### 1.2 VOC MONITORING RESULTS

Field	Units			
Stack ID.		SB1	SB2	SB3
Date		18/02/2009	18/02/2009	18/02/2009
Sample Period (GMT)	From hh:mm	12:05	14:00	15:20
	To hh:mm	12:35	14:30	15:50
Sample Duration	min	30	30	30
30-minute mean Concentration for Period	mg m <sup>-3</sup> , Ref. Conditions	<b>181.1</b>	<b>72.3</b>	<b>28.5</b>
Expanded Uncertainty	mg m <sup>-3</sup> , 95% conf. k=2	+/-11.3	+/-10.7	+/-1.4
Emission Limit Value (ELV)	mg m <sup>-3</sup> , Ref. Conditions	150	150	150
Percentage of ELV	%	121	48	19
Reference Conditions	273.15K, 101.325 kPa, Wet Gas Basis			

**Notes: -**

The VOC (mg/m<sup>3</sup>) results above are expressed as Carbon and these were calculated using the prescribed method described in the Environmental Agency Technical Guidance Note M16 for Volatile Organic Compounds.

# **NATIONAL PHYSICAL LABORATORY**

## **Continuation Sheet**

### **1.3 PLANT AND EQUIPMENT OPERATING INFORMATION**

#### **1.3.1 Paxford Composites Spray Booths 1-3**

The spray booths at Paxford Composites consist of three sealed rooms, approximately the size of a large household car garages. They are used as batch processes and air is pumped into the booths from outside the building and the spray painting process is carried out manually using 2 to 3 skilled workers. The air inside the booth can be heated if required for curing treatment of the components. Owing to the business requirements, many layers of paint are required to achieve the smooth finish. As a result, a typical single layer spray time would normally be approximately 20 minutes but depends on the size and type of item being sprayed. This process is then repeated after the item has been brushed down and ready for a further coating. The paint filled air inside the booths are removed via ceiling filters (which covers the entire surface areas of the ceilings) to remove particulate matter. The air is then exhausted to atmosphere.

On each spray booth, the sample position was downstream of the particulate filters. The sample points were 10mm holes, suitable for monitoring VOCs to the relevant CEN Standard, See Figures 1-3.

At the time of monitoring, spray painting workers were asked to carry out their normal task in order to create representative conditions of routine activity within the spray booths. It should be noted that at the time of monitoring, there were no items being physically sprayed, therefore all of the paint would have been expelled through the exhaust ducting rather than coating surfaces of the items. It was the site's opinion that this may lead to higher levels of VOCs being emitted than would usually be seen.

### **1.4 MONITORING DEVIATIONS**

The testing was fully compliant with the relative standard BS EN13526: 2002. See Appendix 1 for Test Techniques and Protocols.

### **1.5 CONCLUSIONS**

Concentrations of volatile organic compounds, measured from the three spray booths at Paxford Composites Ltd, was carried out on the 18<sup>th</sup> February 2009.

### **1.6 REFERENCES**

1. Guidance on Assessing Measurement Uncertainty in Stack Emissions Monitoring, by Pullen J, Source Testing Association, Quality Guidance Note QGN1 and Q1035-98.

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

<b>Part 2: - Supporting Information-Compliance Report</b>			
Process Operator <b>Paxford Composites Ltd</b>		LA Permit No <b>B01/02</b>	
Address <b>Paxford Composites Ltd</b> <b>2-4 Redwongs Way</b> <b>Huntingdon</b> <b>Cambridgeshire</b> <b>PE29 7HB</b>		Contact <b>Grahame Bloxham</b>	
		Tel No <b>01480 453537</b>	
		Email <a href="mailto:Grahame.bloxham@paxfordcomposites.co.uk">Grahame.bloxham@paxfordcomposites.co.uk</a>	
Tests carried out	<b>Compliance VOCs of 3 Spray Booths</b>	Dates tests carried out <b>18<sup>th</sup> February 2008</b>	
Testing laboratory <b>National Physical Laboratory</b>		UKAS Accreditation No <b>0002</b>	
Address <b>Hampton Road</b> <b>Teddington</b> <b>TW11 OLW</b>		Contact <b>Kevin Blakley</b>	
		Tel No <b>020 8943 6118</b>	
		Email <a href="mailto:kevin.blakley@npl.co.uk">kevin.blakley@npl.co.uk</a>	
Species to be monitored		<b>Volatile Organic Compounds</b>	
Emissions Limit Values (ELV)		<b>Volatile Organic Compounds - 150 mg m<sup>-3</sup> (expressed as carbon)</b>	
Compliance with standards		<b>Yes</b>	
Deviations from standards			<b>No</b>
Corrective actions required			<b>No</b>
Test Team	<b>Kevin Blakley and Robert Elliott</b>		
Full report reference number		<b>PX01FEB09</b>	
Summary report submitted by  <b>Robert Elliott</b>		Signature	MCERTs ID No: -  <b>Level 2, TE1,2,3,4</b>

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**CONTENTS LIST**

	Page
 <b>PART 2: SUPPORTING INFORMATION</b>	
<b>PART 2 COVER SHEET</b>	6
<b>2.1 APPENDIX 1</b>	
2.1.1 Emission Testing Personnel Information	8
2.1.2 Test Techniques and Protocols	8
 <b>2.2 APPENDIX 2</b>	
2.2.1 Sample Points Details	9
2.2.2 VOC 1 Minute Averaged Graphs	11
2.2.3 Gas Measurements	15
2.2.4 Calculations Used in Reporting Results	18
2.2.5 Uncertainty Calculations	19

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.1 APPENDIX 1**

**2.1.1 Emission Testing Personnel Information**

NAME	NPL Position	MCerts I.D. No.	Level/Endorsements	Function
Robert Elliott	HRS	MM-03-319	Level 2, TE1, TE2, TE3, TE4	Team Leader
Kevin Blakley	SRS	MM-03-317	Level 2, TE1, TE2, TE3	Team Leader

**2.1.2 Test Techniques and Protocols**

<b>Date of Tests</b>	18 <sup>th</sup> February 2009
<b>Main Pollutants Measured</b>	1. Volatile Organic Compounds (VOCs)
<b>Test Methods</b>	1. Flame Ionisation Detector, Sick-Maihak. Serial No. AS0202 to BS EN 13526:2002. MCertified Instrument No – Sira MC 040037/02
<b>Procedure</b>	1. <b>QPAS B 538:-</b> STACK GAS SAMPLING FOR CARBON MONOXIDE, CARBON DIOXIDE, OXYGEN, SULPHUR DIOXIDE, VOLATILE ORGANIC COMPOUNDS AND NITROGEN OXIDES USING EXTRACTIVE INSTRUMENTAL TECHNIQUES.



**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**


**2.2 APPENDIX 2**

**2.2.1 Sample Point Details**

**Figure 1. Spray Booth 1 Sample Position**




**Figure 2. Spray Booth 2 Sample Position**



**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**Figure 3. Spray Booth 3 Sample Position**




**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.2.2 VOC 1 Minute Averaged Graphs**

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.2.2 Figure 5. Spray Booth 2. 1-Minute Averages of Volatile Organic Compounds (VOC).  
Expressed as mg m<sup>-3</sup> (Carbon) at 273K, 101.3kPa, Wet Gas Basis**



Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02


Checked by:

Version 1

Page 12 of 23

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

2.2.2 Figure 4. Spray Booth 1. 1-Minute Averages of Volatile Organic Compounds (VOC).  
Expressed as  $\text{mg m}^{-3}$  (Carbon) at 273K, 101.3kPa, Wet Gas Basis



Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02


Checked by:

Version 1

Page 13 of 23

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

2.2.2 Figure 5. Spray Booth 2. 1-Minute Averages of Volatile Organic Compounds (VOC).  
Expressed as  $\text{mg m}^{-3}$  (Carbon) at 273K, 101.3kPa, Wet Gas Basis



Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02

Checked by:

Version 1

Page 14 of 23

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.2.3 Gas Measurements**

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

**Calibration Log**

Instrument SN: <b>AS0202</b>		Type			TEST 12:05-12:40 18/2/09		
Mobile Lab: <b>NPL</b>		VOC			SPRAY BOOTY !		
Gas Cyl. No. <b>WPC 3804N6</b>	Gas Cyl. Conc. <b>51.4</b>	Sport propane/ethylene Delete as app.		Date: <b>18/2/09</b>	Instrument Range: <b>0-100 ppm</b>		
Cal Type	Time (Start and Finish)	Initial Reading	Initial Span/Setting (if app.)	Final Reading	Final Span/Setting (if app.)	System(s) / Pressure (bar/psi) (delete as app.)	Signature
Check Zero	11:20	-0.310	3.72			0	RC
Zero Adj	11:22			0.002	4.32	0	RC
Check Span	11:30					0	RC
Span Adj	11:35	5.346	2.08	5.141	2.80	200	RC
Check Zero	11:38	-0.024	4.32			0	RC
Check Zero	12:42	-0.032				5	RC
Check Span	12:45	5.148				5	RC
Zero Adj							
Span Adj							
Check Zero	12:48	-0.024				5	RC

**IF PRESSURE IS LESS THAN 500 PSI (35 bar) CONTACT SUPPLIER IMMEDIATELY**

NPL Ltd : 29 Jan 2007  
Author: REE

Revision 2.1  
OPAS B 508  
File: \BIB\NPL CA\LOG-NPL\_Version 2.XLS CA-LOG NOV VOC



**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

NPL Ltd : 29 Jan 2007  
Author: REE

FILE: VME.CRY/CALL/DC-NPL Version 2.1.5

**ATmospheric Pressure is LESS THAN 500 PSI (35 bar) CONTACT SUPPLIER IMMEDIATELY**

Instrument SN: <b>AS0202</b>		Type	VOC		SPRAY ROOM 2		14:00-14:30	
Mobile Lab: NPL								
Gas Cyl. No. <b>NPL 1102FANE</b>	Gas Cyl Conc. <b>Si 4</b>	ppm propylene Delete as app.		Date: <b>18/2/04</b>	Instrument Range: <b>0-1000 ppm</b>			
Cyl Type	Time (Start and Finish)	Initial Reading	Final Reading	Initial Gain/Setting (if app.)	Final Gain/Setting (if app.)	System(s) Direct/O	Total Regulator Pressure (bar/Pa) (delete as app.)	Signature
Check Zero		0.043	4.32	-0.000	4.30	0	0.1	RE
Zero Adj						0	19.0	RE
Check Span		8.362	26.0	5.154	24.6	0		RE
Span Adj						0	0.1	RE
Check Zero		-0.003	4.30			5	0.1	RE
Check Zero		14.40	0.030			5	0.1	RE
Check Span		12.43	5.095			5	0.1	RE
Zero Adj						5	0.1	RE
Span Adj						5	0.1	RE
Check Zero		14.48	0.305			5	0.1	RE

## Calibration Log

~~IF REGULAR LABOR PRESSURE IS LESS THAN 500 PSI (35 bar) CONTACT SUPPLIER IMMEDIATELY~~

Page 18 of 23

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.2.4 Calculations Used in Reporting Results**

The following equation can be used to convert propane volume concentrations to total organic carbon mass concentrations:

$$C_m = C_v \frac{3 \times M_c}{V_m} \text{ mg/m}^3$$

where:-

$C_m$  is the TOC concentration in milligrams per cubic metre (273 K;1013 hPa).

$C_v$  is the volume concentration of propane in ppm (by volume).

$M_c$  is the molar mass of carbon (=12 g/mole).

$V_m$  is the molar volume (=22,4 l/mole).

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**2.2.5 Uncertainty Calculations**

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 1

Limit value	150	mg.m <sup>-3</sup> (corrected) Carbon	Gas	Propane
Measured concentration	112.83	ppm	Full Scale	1000
Measured concentration	181.10	mg.m <sup>-3</sup> (273K, 101.3kPa) Carbon	Cal gas conc	51.4
			Conversion	0.62305296
			Full Scale	623.0529595
			Cal gas conc	32.02492212
				mg.m <sup>-3</sup> (Carbon)
				mg.m <sup>-3</sup> (Carbon)

Correction for reference conditions					
	ref	O2, %	Moisture, %	Pressure, KPa	Temperature, K
		21.00	0.00	101.30	273.00
	measured	21.00	0.00	101.30	273.00
Factors		1.00	1.00	1.00	1.00
Correction Factor		1.00			

Performance characteristics	Value		specification
Response time	10	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0	% of value/kPa	0.750
ambient temperature dependence	0.025	% full scale/10K	0.300
NH3 (20 mg/m3)	0	mg/m3	
CO2 (15%)	0	% by vol	
H2O (30%)	0.0	% by vol	4.000
dependence on voltage	1.05	% full scale/10V	2%fs/10V
losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m3
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH3 range	0	0	0
CO2 range	0	15	0
H2O range	0	0	0
Instrument Voltage Rating			110
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions					
Performance characteristic		Uncertainty		Value of uncertainty quantity	
Standard deviation of repeatability at zero		U <sub>io</sub>		for mean	use rep at span
Standard deviation of repeatability at span level		U <sub>rs</sub>		for mean	0.02
Lack of fit		U <sub>fit</sub>			-5.25
Drift		U <sub>odr</sub>			0.00
volume or pressure flow dependence		U <sub>spres</sub>			0.00
atmospheric pressure dependence		U <sub>spres</sub>			0.00
ambient temperature dependence		U <sub>temp</sub>			0.00
NH3 (20 mg/m3)		U <sub>interf</sub>			0.00
CO2 (15%)					0.00
H2O (30%)					0.00
Dependence on voltage		U <sub>volt</sub>			0.33
losses in the line (leak)		U <sub>leak</sub>			0.00
Uncertainty of calibration gas		U <sub>calib</sub>			2.09

Use largest negative or positive interferent effect	
0	0.00
0	0.00
0	0.00
0	0.00
Interference uncertainty	0.00

Measurement uncertainty		Result	181.10	mg/m <sup>3</sup>
Combined uncertainty			5.66	mg/m <sup>3</sup>
Expanded uncertainty	k =	2	11.33	mg/m <sup>3</sup>
Uncertainty corrected to std conds			11.33	mg.m-3 (corrected)
Expanded uncertainty	expressed with a level of confidence of 95%		7.55 %ELV	
Expanded uncertainty	expressed with a level of confidence of 95%		11.33 mg.m <sup>-3</sup>	

Note: Enter values into green boxes  
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R Robinson, NPL

Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02

Checked by:

Version 1

Page 21 of 23

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 2

Limit value	150	mg.m <sup>-3</sup> (corrected) Carbon	Gas	Propane
Measured concentration	45.05	ppm	Full Scale	1000
Measured concentration	72.30	mg.m <sup>-3</sup> (273K, 101.3kPa) Carbon	Cal gas conc	51.4
			Conversion	0.62305296
			Full Scale	623.0529595
			Cal gas conc	32.02492212
				mg.m <sup>-3</sup> (Carbon)
				mg.m <sup>-3</sup> (Carbon)

Correction for reference conditions				
	O <sub>2</sub> , %	Moisture, %	Pressure, KPa	Temperature, K
ref	21.00	0.00	101.30	273.00
measured	21.00	0.00	101.30	273.00
Factors	1.00	1.00	1.00	1.00
Correction Factor	1.00			

Performance characteristics	Value		specification
Response time	10	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
atmospheric pressure dependence	0	% of value/kPa	0.750
ambient temperature dependence	0.025	% full scale/10K	0.300
NH <sub>3</sub> (20 mg/m <sup>3</sup> )	0	mg/m <sup>3</sup>	
CO <sub>2</sub> (15%)	0	% by vol	
H <sub>2</sub> O (30%)	0.0	% by vol	4.000
dependence on voltage	1.05	% full scale/10V	2%fs/10V
losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m <sup>3</sup>
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH <sub>3</sub> range	0	0	0
CO <sub>2</sub> range	0	15	0
H <sub>2</sub> O range	0	0	0
Instrument Voltage Rating			110
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions					
Performance characteristic		Uncertainty		Value of uncertainty quantity	
Standard deviation of repeatability at zero		u <sub>0</sub>		for mean	use rep at span
Standard deviation of repeatability at span level		u <sub>s</sub>		for mean	0.02
Lack of fit		u <sub>fit</sub>			-5.25
Drift		u <sub>dr</sub>			0.00
volume or pressure flow dependence		u <sub>spres</sub>			0.00
atmospheric pressure dependence		u <sub>apres</sub>			0.00
ambient temperature dependence		u <sub>tomp</sub>			0.00
NH <sub>3</sub> (20 mg/m <sup>3</sup> )		u <sub>interf</sub>			0.00
CO <sub>2</sub> (15%)					0.00
H <sub>2</sub> O (30%)					0.00
Dependence on voltage		u <sub>vot</sub>			0.33
losses in the line (leak)		u <sub>leak</sub>			0.00
Uncertainty of calibration gas		u <sub>calib</sub>			0.83

Use largest negative or positive interferent effect	
0	0.00
0	0.00
0	0.00
0	0.00
Interference uncertainty	0.00

Measurement uncertainty	Result	72.30	mg/m <sup>3</sup>
Combined uncertainty		5.33	mg/m <sup>3</sup>
Expanded uncertainty	k = 2	10.66	mg/m <sup>3</sup>
Uncertainty corrected to std conds		10.66	mg.m-3(corrected)
Expanded uncertainty	expressed with a level of confidence of 95%	7.10	% ELV
Expanded uncertainty	expressed with a level of confidence of 95%	10.66	mg.m <sup>-3</sup>

Note: Enter values into green boxes  
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R Robinson, NPL

Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02

Checked by:

Version 1

Page 22 of 23

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### Uncertainty calculation for Gaseous Measurement BS EN 13526 VOCs

SPRAY BOOTH 3

Limit value	150	mg.m <sup>-3</sup> (corrected) Carbon	Gas	Propane
Measured concentration	17.76	ppm	Full Scale	100
Measured concentration	28.50	mg.m <sup>-3</sup> (273K, 101.3kPa) Carbon	Cal gas conc	51.4
			Conversion	0.62305296
			Full Scale	62.30529595
			Cal gas conc	32.02492212
				mg.m <sup>-3</sup> (Carbon)
				mg.m <sup>-3</sup> (Carbon)

Correction for reference conditions				
	O <sub>2</sub> , %	Moisture, %	Pressure, kPa	Temperature, K
ref	21.00	0.00	101.30	273.00
measured	21.00	0.00	101.30	273.00
Factors	1.00	1.00	1.00	1.00
Correction Factor	1.00			

Performance characteristics	Value		specification
Response time	10	seconds	180.000
Number of readings in measurement	30		
Repeatability at zero	0.005	% full scale	0.200
Repeatability at span level	0.083	% full scale	2.000
Deviation from linearity	-1.46	% of value	2.000
Zero drift	0	% full scale	2.000
Span drift	0	% full scale	2.000
Volume or pressure flow dependence	0.02	% of full scale/kPa	0.033
Atmospheric pressure dependence	0	% of value/kPa	0.750
Ambient temperature dependence	0.025	% full scale/10K	0.300
NH <sub>3</sub> (20 mg/m <sup>3</sup> )	0	mg/m <sup>3</sup>	
CO <sub>2</sub> (15%)	0	% by vol	
H <sub>2</sub> O (30%)	0.0	% by vol	4.000
Dependence on voltage	1.05	% full scale/10V	2%fs/10V
Losses in the line (leak)	0	% of value	2% of value
Uncertainty of calibration gas	2	% of value	

Effect of drift
0.00 mg/m <sup>3</sup>
0.00 % value

	ranges		
	min	max	value at calib
flow	1.9	2.1	2
pressure	101.30	101.3	101.3
temp	289	289	289
NH <sub>3</sub> range	0	0	0
CO <sub>2</sub> range	0	15	0
H <sub>2</sub> O range	0	0	0
Instrument Voltage Rating			110
Voltage	104.5	115.5	110

Measurement performance related to stationary conditions					
Performance characteristic		Uncertainty		Value of uncertainty quantity	
Standard deviation of repeatability at zero		u <sub>0</sub>		for mean	use rep at span
Standard deviation of repeatability at span level		u <sub>s</sub>		for mean	0.02
Lack of fit		u <sub>fit</sub>			-0.53
Drift		u <sub>dr</sub>			0.00
Volume or pressure flow dependence		u <sub>spres</sub>			0.00
Atmospheric pressure dependence		u <sub>apres</sub>			0.00
Ambient temperature dependence		u <sub>tomp</sub>			0.00
NH <sub>3</sub> (20 mg/m <sup>3</sup> )		u <sub>interf</sub>			0.00
CO <sub>2</sub> (15%)					0.00
H <sub>2</sub> O (30%)					0.00
Dependence on voltage		u <sub>vot</sub>			0.33
Losses in the line (leak)		u <sub>leak</sub>			0.00
Uncertainty of calibration gas		u <sub>calib</sub>			0.33

Use largest negative or positive interferent effect	
0	0.00
0	0.00
0	0.00
0	0.00
Interference uncertainty	0.00

Measurement uncertainty	Result	28.50	mg/m <sup>3</sup>
Combined uncertainty		0.70	mg/m <sup>3</sup>
Expanded uncertainty	k = 2	1.41	mg/m <sup>3</sup>
Expanded uncertainty		1.41	mg.m <sup>-3</sup> (corrected)
Expanded uncertainty	expressed with a level of confidence of 95%	0.94	% ELV
Expanded uncertainty	expressed with a level of confidence of 95%	1.41	mg.m <sup>-3</sup>

Note: Enter values into green boxes  
Dark blue boxes indicate information that can be obtained from MCERTS tests

Developed by R Robinson, NPL

Reference: 105435/QE8400/PX01FEB09

Permit No: B01/02

Checked by:

Version 1

Page 23 of 23